



**Australian Government**  
**Department of Health and Aged Care**  
Therapeutic Goods Administration

## Public Summary

**Summary for ARTG Entry:** 147088 Aus Inno Pty Ltd - Chamber, patient, hyperbaric

**ARTG entry for** Medical Device Included Class IIb  
**Sponsor** Aus Inno Pty Ltd  
**Postal Address** 411/1-3 Sturt Place, St Ives, NSW, 2075  
Australia  
**ARTG Start Date** 7/11/2007  
**Product Category** Medical Device Class IIb  
**Status** Active  
**Approval Area** Medical Devices

### Conditions

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.  
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

### Manufacturers

Name	Address
Perry Baromedical Corporation	3750 Prospect Avenue Riviera Beach, FL, 33404 United States Of America

### Products

#### 1 . Chamber, patient, hyperbaric

<b>Product Type</b>	Single Device Product	<b>Effective Date</b>	28/04/2023
<b>GMDN</b>	12061 Chamber, patient, hyperbaric		
<b>Intended Purpose</b>	The monoplace hyperbaric chamber system provides non-invasive hyperbaric 100% oxygen therapy to an operating pressure of 303.9kPa to treat acute and chronic medical conditions.		

### Specific Conditions

No Specific Conditions included on Record

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/website-copyright.htm>.

Public Summary