



**Australian Government**  
**Department of Health and Aged Care**  
Therapeutic Goods Administration

## Public Summary

<b>Summary for ARTG Entry:</b>	343493	Connexion Surgical Pty Ltd - MagnetOs Granules - Bone matrix implant, synthetic
<b>ARTG entry for</b>	Medical Device Included Class III	
<b>Sponsor</b>	Connexion Surgical Pty Ltd	
<b>Postal Address</b>	12 Finch Avenue, Concord, NSW, 2137 Australia	
<b>ARTG Start Date</b>	15/09/2020	
<b>Product Category</b>	Medical Device Class III	
<b>Status</b>	Active	
<b>Approval Area</b>	Medical Devices	

### Conditions

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.  
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

### Manufacturers

Name	Address
Kuros Biosciences BV	Professor Bronkhorstlaan 10 , MB Bilthoven, 3723 Netherlands

### Products

#### 1 . MagnetOs Granules - Bone matrix implant, synthetic

<b>Product Type</b>	Single Device Product	<b>Effective Date</b>	15/09/2020
<b>GMDN</b>	17751 Bone matrix implant, synthetic		
<b>Functional Description</b>	MagnetOs is a synthetic, Tri-Calcium Phosphate and Hydroxyapatite resorbable micro-structured bone void filler for the repair of bony defects. MagnetOs has a porous trabecular structure that resembles the interconnected porosity of human cancellous bone.		
<b>Intended Purpose</b>	MagnetOs is intended for use as bone void filler for voids and gaps that are not intrinsic to the stability of the bony structure. MagnetOs is indicated for use in the treatment of surgically created osseous defects or osseous defects resulting from traumatic injury to the bone. MagnetOs is intended to be packed into bony voids or gaps of the skeletal system (i.e. extremities, spine, cranial, mandible, maxilla and pelvis) and may be combined with autogenous bone, blood, PRP and/or bone marrow.		
<b>Variant information</b>	Volume (mL) 0.5-20 Size 150-4000um		

### Specific Conditions

No Specific Conditions included on Record

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/website-copyright.htm>.

Public Summary